



UNIFIED FIRE AUTHORITY

MEMORANDUM

24-073

May 30, 2024

TO: All Personnel

FROM: Division Chief Ayres

SUBJECT: ZOLL Monitor ETCO2 Alarm Configuration

At the request of our Medical Director, BioTech will be updating the configuration of our Zoll monitors so that the low ETCO2 alarm activates if measured ETCO2 drops below 20.

For questions related to ETCO2 monitoring please contact EMS Training Specialist Amanda Lawrence (alawrence@unifiedfire.org).

For questions related to Zoll cardiac monitor function and technical operation, please contact Staff Captain Andrus in BioTech (rjandrus@unifiedfire.org).